

**How to Cite:**

Singh, D. (2021). Cone-beam computed tomography: A new era in clinical orthodontics. *International Journal of Health Sciences*, 5(S1), 411–427. <https://doi.org/10.53730/ijhs.v5nS1.5677>

# Cone-beam computed tomography: A new era in clinical orthodontics

**Disha Singh**

PG Student 1<sup>st</sup> Year, Department of Orthodontics, Desh Bhagat Dental College & Hospital, Mandi Gobindgarh, Punjab, India

Email: [dr.dishasingh@icloud.com](mailto:dr.dishasingh@icloud.com)

**Abstract**---Cone beam computed tomography (CBCT) is an important source of three-dimensional volumetric data in clinical orthodontics. Due to the progress in the technology of CBCT, for orthodontic clinical diagnosis, treatment and follow-up, it supply much more reliable information compared to conventional radiography. But it carries both risks and benefits in orthodontics and the principal risks and limitations include ionizing radiation, the presence of artifacts, higher cost, limited accessibility, and the need for additional training. However, this imaging modality has several recognized indications in orthodontics, such as the assessment of impacted and ectopic teeth, assessment of pharyngeal airway, assessment of mini-implant sites, evaluation of craniofacial abnormalities, evaluation of sinus anatomy or pathology, evaluation of root resorption, evaluation of the cortical bone plate, and orthognathic surgery planning and evaluation. CBCT is particularly justified when it brings a benefit to the patient or changes the outcome of the treatment when compared with conventional imaging techniques. Therefore, CBCT should be considered for clinical orthodontics for selected patients. this article presents discussion on radiation dosage of CBCT and other imaging techniques used in orthodontics, advantages and disadvantages of CBCT in orthodontics, justifying the use of CBCT in orthodontics, and the benefits and evidence-based indications of CBCT in orthodontics.

**Keywords**---CBCT, cone-beam CT, CBCT review.

**Introduction**

Cone-beam computed tomography (CBCT) is a radiographic technique introduced to the United States dental market in 2001. Since the discovery of the X-ray more than a century ago, few other diagnostic imaging modalities have impacted dental practice to the extent that CBCT has. Since CBCT introduction, the progress

---

International Journal of Health Sciences ISSN 2550-6978 E-ISSN 2550-696X © 2021.

**Corresponding author:** Singh, D.; Email: [dr.dishasingh@icloud.com](mailto:dr.dishasingh@icloud.com)

Manuscript submitted: 18 April 2021, Manuscript revised: 9 June 2021, Accepted for publication: 5 July 2021

made in CBCT maxillofacial applications has been remarkable. CBCT technology uses a cone-shaped source of ionizing radiation and a two-dimensional detector. It provides multidimensional and dimensionally accurate images for diagnosis and treatment planning. These images contain isotropic voxels (volume elements) such that each volume element has equal dimensions in all three orthogonal planes, allowing accurate multi-planar images in any direction desired by the practitioner.

CBCT has attracted significant attention from practitioners who seek to enhance diagnosis and treatment for their patients [2]. Indications of CBCT in orthodontics have been documented. However, risks and limitations of CBCT need to be explored and weighed against the benefits of CBCT in each case. Practitioners of the healing arts must minimize harm to their patients. It is therefore necessary to find valid and robust evidence on which to base the selection of CBCT imaging for the orthodontic patient. There is a vast literature on CBCT in orthodontics, including several opposing views. Therefore, the objective of this narrative review is to answer this question: when does CBCT add value to the practice of orthodontics? To answer this question, this article presents discussion on radiation dosage of CBCT in orthodontics, limitations of CBCT in orthodontics, justifying the use of CBCT in orthodontics, and the benefits and evidence-based indications of CBCT in orthodontics. Ten orthodontic cases in which CBCT was utilized are presented. Understanding the indications for CBCT in orthodontics and weighing its risks and benefits allow the orthodontist to be able to prescribe CBCT when it brings value to the orthodontic patient.

### **Radiation Dosage of CBCT in Orthodontics**

Theoretically, any amount of ionizing radiation, no matter how small, has the potential to cause a deleterious effect. Radiation is a carcinogen, and current radiation protection protocols are based upon the linear non-threshold (LNT) assumption that even very low doses of radiation can cause cancer. Most patients who undergo orthodontic therapy are children[4], and children of orthodontic age are radiosensitive and susceptible to the untoward effects of ionizing radiation[56], whereas adults are more resistant. Children have higher risk from ionizing radiation for two reasons: they have higher cell and tissue sensitivity to radiation than adults, and they have a longer lifespan than adults in which radiation-induced changes may manifest.[7,12]

Radiation carcinogenesis has a stochastic effect which means that the probability of cancer increases with increased dose, but the severity of cancer is not related to the dose [3]. For instance, a similar malignancy developed later in life can be caused by any radiation dosage, but the chance of its occurrence increases with a higher dose. Generally, children's exposure to low radiation doses has the effect of a small yet insignificant increase in the risk of a fatal cancer that may develop during life [13,14]. In addition to the age factor, the risk of cancer arising from radiation varies depending on gender, exposure type (acute or chronic), and radiation type. For instance, female patients are slightly more radiosensitive than male patients [15]. In other words, not all radiation exposures have the same effect. There are differences between dental and medical radiographic imaging. In the medical field, medical computed tomography (CT) scans carry the highest

risk, and the risk assessment has shown that these scans have become a leading source of future risk to the general population [16]. To estimate the risk of ionizing radiation, the effective dose concept is used. To compute the effective dose, the total amount of absorbed dosages by the tissues is multiplied by the tissue weighting factors [17].

CBCT effective doses are smaller than those of medical CT [18–27]. However, there is a wide range of effective doses that are present across different CBCT machines. This large range of effective doses is strongly correlated with the size of the field of view (FOV) [28]. If the FOV of CBCT is increased, the effective dose increases as well [29]. Reducing the size of the FOV is therefore one of the greatest and easiest ways to reduce the effective dose of CBCT. In addition, reducing the scan time, number of projections, and the MAs (Milliampereseconds) has an additional role in reducing the dose as well [30,31]. In fact, small and strategic adjustments in exposure parameters can result in significant reduction in the effective dose without significantly compromising the image quality [32,33]. However, significant modifications of these parameters aimed at significant reduction of the effective dose can reduce the image quality [28,34]. Therefore, clinical judgment should be exercised with dose reduction efforts in order to maintain diagnostic and quality images.

Because the dose received is strongly related to the field size, a small FOV can be selected for the region of interest that triggers the interest in CBCT acquisition [35]. In order to optimize the use of CBCT, the FOV should be justifiable, patient-specific, and indication-oriented [36]. An impacted canine, for example, would not require a large volume CBCT scan. A small CBCT volume of  $40 \times 40$  may be sufficient, patient-specific, and indication-oriented. Using smaller volumes benefits the patient because it can reduce the effective dose [28,37–40]. In addition, it benefits the practitioner, because small CBCT volumes do not include areas in the head that are difficult for most dental practitioners to interpret, and thus reduces time spent on radiographic interpretation [41].

Whereas effective doses of CBCT are less than those of medical CT, CBCT dosages are generally higher than effective doses of panoramic and cephalometric imaging. The effective dose of a digital panoramic radiograph has the range of 6–38 micro Sieverts ( $\mu\text{Sv}$ ) [29,42–47], and the effective dose of a cephalometric radiograph has the range of 2–10  $\mu\text{Sv}$  [23,46,48]. On the other hand, the range of effective doses of CBCT is very large and has been reported to be 5.3–1025  $\mu\text{Sv}$ , depending on the size of the FOV, specific technique factors, and the machine itself [25,29,34,37,42–47,49–54]. One legacy CBCT machine had a large field-of-view setting in which the effective dose exceeded 1000  $\mu\text{Sv}$  [52]. To put this in perspective, the effective dose of a medical CT for the head is approximately 1000–2000  $\mu\text{Sv}$  [26]. It must be stated that most of the current CBCT dosages are in the lower half of the reported range, and significant efforts are being made to standardize different CBCT scanners and to further reduce CBCT dosages to the point that they are close to the panoramic and cephalometric radiographic dosages [55]. As Table 1 demonstrates, the combined panoramic and cephalometric radiographic dosages and the lowest CBCT dosage for some machines and significantly reduced exposure settings (i.e., FOV, mAs, scan time) may actually overlap. Some CBCT machines have the capability of reducing the

amount of radiation dose for different patient sizes while maintaining optimal image detail and quality. Moreover, a new technology called the Dose Reduction Technology (DRT) can allow the clinician to set the machine in the DRT mode, which results in low dosages that rival two-dimensional imaging such as panoramic radiography.

Table 1. Comparison between the effective dose of digital panoramic radiography, cephalometric radiography, cone-beam computed tomography (CBCT), and medical computed tomography (CT).

Imaging Technique	Range of Effective Dose ( $\mu\text{Sv}$ ) Reported in the Literature
Panoramic radiography	6–38
Cephalometric radiography	2–10
CBCT	5.3–1025
Medical head CT	1000–2000

Besides the large range of CBCT reported doses, these values may in fact differ across different ages. For instance, children have higher effective doses because they are smaller than adults [56]. The difference in size between children and adults results in the higher proximity of radiosensitive organs (e.g., thyroid gland) in children to the FOV, which results in a larger effective dose for children [46]. This occurs even if the exposure protocols are exactly the same. Therefore, the cancer risk per unit of radiation dose is higher for children than for adults [57].

Collective effective dose, measured in person-Sv, is another concept in radiation biology. It is defined as the product of the effective dose and the number of individuals exposed. This concept is frequently mentioned in medical imaging because CT scans have high dosages. The collective medical effective dose in the United States of a population of about 300 million was estimated to be 900,000 person-Sv in 2006. This figure is about seven-fold the estimate made in 1982 (124,000 person-Sv), due to the increased popularity of CT scans and nuclear medicine [58]. These two modalities account for 75% of the collective medical effective dose [59], and it is estimated that approximately 1.5 to 2% of all the cancers developing in the U.S. are due to the use of CT alone [60].

This may be a public health issue, but it is related mainly to CT scans. As previously mentioned, and as Table 1 demonstrates, CBCT radiation doses are fortunately lower than the corresponding doses for medical CT. Yet there is one resemblance that can be observed here; the increased popularity of CBCT in orthodontics over time will inevitably result in the increase of collective effective dose for orthodontic patients, thus increasing the likelihood of radiation risks in these patients [61,62]. Because children are sensitive to radiation, the use of thyroid protection (lead apron with collar) has been recommended [63]. Lead shielding significantly reduces the effective dose and is generally an effective way to reduce the risks of ionizing radiation [53].

### **Limitations and Liability Associated with the Use of CBCT in Orthodontics**

Besides the exposure to ionizing radiation, CBCT comes with other limitations and concerns. For example, CBCT scanners have higher cost and limited accessibility when compared to conventional radiographic imaging techniques. In addition, CBCT images are sufficient for visualization of teeth and bone, but are unable to represent the internal structure of soft tissues or soft tissue lesions with high accuracy [64,65]. Inherent artifacts that may be present in CBCT images include beam hardening [66]. In general, metal artifacts are observed on CBCT images in the vicinity of metals [67]. In orthodontics, these artifacts can be noted on the images around orthodontic brackets and bands (scattering) [68].

Also, CBCT images can display noise, cupping artifacts, or scatter [69]. It is possible to acquire CBCT during orthodontic treatment, but the images may include beam hardening and scatter around orthodontic appliances. Other limitations may include motion artifacts, especially in young orthodontic patients who are more likely to move during long CBCT scans [70]. These limitations inherent to CBCT should be considered because they can affect the image quality. CBCT image quality is not comparable across different scanners [71]. There are approximately 50 commercially available CBCT models and scanners with variable image quality. Clinicians who are unfamiliar with CBCT image quality may not be able to compare different scanners in regard to their images. While CBCT images are considered accurate and reliable in terms of linear measurements [72–78], CBCT images may occasionally present false positives and false negatives. For example, CBCT images may not produce a reliable presentation of a thin cortical bone [79]. Misinterpretation of CBCT images may affect orthodontic decision making. Further, an artifact may be confused with the presence of pathology and may therefore lead to false diagnoses.

Presentation of CBCT images through volume rendering or Maximum Intensity Projection (MIP) may increase the likelihood of false findings. These illustrations are created based on sophisticated software algorithms, and therefore they may not always be accurate. Therefore, evaluation of the volume through axial, sagittal, and coronal views is required. Such evaluation is technically demanding and may be difficult initially for some practitioners. Interpretation of CBCT scans requires skills and knowledge beyond that obtained at dental school [80,81].

Finally, with the use of CBCT scanning, the orthodontist bears legal responsibility to report any pathology in the scan [82,83]. There has been significant controversy regarding the orthodontist's liability to report any pathology evident in the scan. As with any radiographic interpretation, the orthodontist is responsible for interpretation of the CBCT volume in its entirety [84]. In some countries, such as the United States, the full interpretation of CBCT is a legal requirement [82,83,85–87]. Some clinicians may choose to refer to an oral and maxillofacial radiologist to transfer these risks [88], and at the same time provide their patients with a specialty level care for the radiographic interpretation of their CBCT scans [89]. When several of these risks and limitations inherent to CBCT imaging are mitigated or eliminated, CBCT becomes an excellent tool to enhance orthodontic diagnosis and treatment planning [89], however, the use of CBCT must be justified according to established guidelines.

### **Justifying the Use of CBCT in Orthodontics According to Established Guidelines**

In orthodontics, the same set of radiographs should not be routinely made for all patients [90,91]. Orthodontists find the panoramic and cephalometric radiography to be sufficient for most initial, progress, and final records [87,92]. However, CBCT may prove to be advantageous in some clinical encounters. The great advantage of CBCT is that it provides images of various dental, oral, and maxillofacial structures in multiple orthogonal images (i.e., coronal, sagittal, axial). CBCT can also provide curved or flat slices of variable thickness. In addition, CBCT provides multi-planar reformatted images, volume rendering, maximum intensity projection, and other 3D visual representations.

Orthodontists and dental practitioners should carefully consider any radiographic examination before it is ordered. This process is called image selection or the use of selection criteria. The selection of CBCT in general is based on the patient's presentation and the need to diagnose, monitor, or evaluate the outcome of a treatment [93]. For any case, the orthodontist should be able to justify the use of CBCT. CBCT can be justified if conventional imaging techniques such as panoramic and cephalometric radiographs fail to provide correct diagnosis or when CBCT has a positive effect on treatment options or treatment optimization [94, 95]. It need not be considered a standard method of diagnosis in orthodontics because conventional two-dimensional radiographic techniques (e.g., panoramic and cephalometric radiographs) usually suffice for orthodontic diagnosis and treatment planning.

Because the concerns about radiation risks are heightened for children, who comprise most orthodontic patients, several position statements have been made by respected organizations. Position statements and clinical guidelines made by reputable international health care organizations are authoritative and defensible. They are released after exhaustive review and appraisal of the literature. The Swiss Association of Dento maxillofacial Radiology recommends that CBCT in orthodontics be used only if it brings additional information compared to conventional two-dimensional imaging [96]. The DIMITRA (Dento maxillofacial pediatric imaging: an investigation towards low-dose radiation induced risks), a European multicenter and a multidisciplinary project, released a position statement encouraging practitioners to follow the principle of ALADAIP—keeping radiation As Low as Diagnostically Acceptable being Indication-oriented and Patient-specific [36]. The clinically relevant ALADAIP directive is especially relevant for young orthodontic patients.

Not a single organization recommended CBCT for all orthodontic patients. For example, the American Dental Association recommended that CBCT be prescribed only when there is an expected diagnostic benefit for the patient or significant improvement in the clinical outcome [93]. The American Academy of Oral and Maxillofacial Radiology recommended the use of CBCT imaging in orthodontics only when there is justification made on an individual basis according to the clinical presentation [85]. The British Orthodontic Society guidelines are comparable, and did not recommend CBCT imaging for all orthodontic patients [10]. Therefore, the strongest theme in these recommendations regarding

prescription of CBCT in orthodontics is that CBCT must be justified on a case-by-case basis and when it has the potential to improve diagnosis or treatment. Prescribing CBCT for all orthodontic patients may be considered a flawed and questionable practice [97].

Despite robust justification of CBCT in selected cases, some authors found insignificant differences in treatment planning decisions when CBCT was used versus conventional imaging [98], and others have stated that, even though CBCT may alter treatment planning, it does not necessarily improve or change orthodontic treatment outcome [99–101]. It is difficult to assess the exact value of CBCT with regards to changing the orthodontic treatment outcome because the evidence on CBCT efficacy and diagnostic value is not obtained from randomized controlled trials, but rather mostly from observational studies or studies with variable hierarchy of evidence [102,103].

### **Benefits and Evidence-Based Indications of CBCT in Orthodontics**

CBCT brings specific and unique diagnostic benefits in orthodontics [104]. The most common indication for CBCT in orthodontics is the 3D assessment of anomalies in dental position such as impactions and ectopic teeth [94]. CBCT allows the visualization of impacted teeth in three dimensions, as well as the evaluation of roots of the impacted and adjacent teeth. It has been suggested that in cases with impacted maxillary canines, CBCT can actually alter treatment planning decisions]. This is due to the fact that conventional panoramic or intraoral radiography may not provide a good assessment of the root status of adjacent teeth, but with CBCT this can be done effectively .This is especially true in cases with severe displacement of the impacted tooth in which an accurate assessment of the impacted and adjacent teeth is essential .Justification of CBCT in these cases increases given that CBCT brings significant value to diagnosis and treatment planning.

In addition to the assessment of anomalies in dental position, CBCT provides information on the stage of dental development, and position and size of the tooth or follicle. CBCT can also provide a great tool for evaluation and detection of any supernumerary teeth . Patients with dentofacial abnormalities and deformities can benefit from CBCT [.For example, CBCT can be prescribed for patients with facial asymmetry, cleft palate, or obstructive sleep apnea [94]. Because structures such as cleft palate and oropharyngeal airway are three-dimensional, it is advantageous to use CBCT for the evaluation of these structures .CBCT also provides three-dimensional assessment for alveolar boundary conditions, craniofacial anatomy, and maxillary transverse dimensions]. CBCT can be used in craniofacial orthodontics in which effects of maxillary expansion, evaluation of the clefts, and the skeletal and soft tissues can be assessed in all dimensions Incidental findings or pathologies discovered via 2D imaging, such as panoramic radiograph, can be better visualized via CBCT. This is especially valuable if the orthodontist desires to evaluate the pathology in three-dimensions and its relationship to the teeth.

If temporary anchorage devices such as mini-implants or mini-plates are planned before or during orthodontic treatment, CBCT can help the practitioner in

evaluating the proposed site for insertion or the status of the temporary anchorage device after the insertion. If the evaluation of the temporomandibular joints (TMJs) is required, CBCT has the potential to provide information about the bony component of the TMJs. CBCT provides better evaluation of the shape and volume of the TMJ condyles when compared to panoramic radiography [94]. However, the articular disk and muscles cannot be visualized via CBCT [70]. These structures are well visualized through magnetic resonance imaging (MRI).

Unlike 2D superimpositions provided by conventional cephalometric radiography, CBCT can provide the clinician with sophisticated 3D superimpositions and treatment assessment when necessary. Assessment of orthognathic surgery can be made via these superimpositions. In addition, assessment of soft tissue changes of the face in orthognathic surgery cases can be made [Whereas CBCT can be used for evaluation of orthodontic surgical cases, the use of CBCT in these cases does not necessarily alter treatment outcome [65]. One of the great features of CBCT is its ability to construct different views, such as a panoramic view of the teeth and adjacent structures and another cephalometric view. Therefore, if a large volume CBCT is made, these views can be generally made without taking additional 2D panoramic and cephalometric radiographs. These images can be reconstructed from the CBCT volume, provided that it includes all areas of interest. Several studies confirmed that the cephalometric view synthesized from CBCT volume is equivalent to the conventional cephalometric radiograph in terms of landmark identification, cephalometric analyses, and the overall diagnostic value

Unlike conventional panoramic imaging (commonly known as the panorex image), CBCT synthesized panoramic views have the advantage of eliminating magnification, ghost images, distortion, and overlaps. However, creating a panoramic view from the CBCT volume should be made with caution in order to obtain a proper and reliable image. The focal trough can be controlled with CBCT synthesized panoramic radiography, whereby it can be modified and customized to the individual's jaw size. For example, it can be increased in the anterior region if the patient has bimaxillary dentoalveolar protrusion, or it can be modified in shape if any impacted or ectopic teeth are present. This results in visualization of objects that would otherwise be located outside the focal trough in conventional panoramic radiography. Finally, the size of the focal trough itself can be decreased or increased. For example, if a practitioner uses a focal trough of 20 mm in width for most cases, the focal trough can be increased to 30 mm in a case of bimaxillary dentoalveolar protrusion in which the teeth are proclined. The ability to change the size of the focal trough in this case results in inclusion of the full length of both maxillary and mandibular incisors in the focal trough.

### **Following the ALARA and ALADAIP Principles**

Practitioners should always follow the basic ALARA directive in radiation protection, keeping radiation "As Low As Reasonably Achievable." A more evolved and specific directive in radiation protection is the ALADAIP principle [36]. It requires practitioners to keep radiation As Low As Diagnostically Acceptable being Indication-oriented and Patient-specific.



The ubiquitous and erroneous practice of taking a large volume CBCT for the whole head merely to synthesize panoramic and cephalometric views does not follow the ALADAIP directive, because it does not keep radiation as low as diagnostically acceptable, and it is neither indication-oriented nor patient-specific. If the orthodontic patient requires only two-dimensional panoramic and cephalometric radiographs, these radiographs could be taken without the additional exposure burden that comes with large CBCT volumes. It also behooves the practitioner to utilize all 3D capabilities of the CBCT scan, and not to be limited to the two-dimensional panoramic and cephalometric views if a large volume is taken.

Whereas panoramic and cephalometric radiographs may not suffice for specific diagnostic tasks, intraoral radiography may be considered in lieu of CBCT imaging. For example, periapical radiographs may suffice for specific diagnostic tasks, such as assessment of root shapes or root resorption or fracture or the evaluation of periodontal status. In other words, if panoramic and cephalometric radiographs are insufficient for these diagnostic tasks, the orthodontist could consider periapical radiography instead of considering CBCT.

When all conventional radiographic techniques are insufficient for diagnosis and treatment, and the orthodontic patient will benefit from CBCT, the clinician should not hesitate to order this imaging technique. If there is a diagnostic benefit to the patient from CBCT in terms of diagnosis and treatment planning, then this benefit outweighs the risks involved. Some patients can benefit dramatically from images provided by CBCT]. Therefore, the orthodontist should not hesitate to order a CBCT scan if certain diagnostic information is needed, particularly if this information cannot be obtained via conventional imaging. However, the scan should always be customized to the patient's needs whenever possible, including the customization of the FOV and other exposure settings in order to reduce and optimize the patient's ionizing radiation exposure.

## **Conclusion**

Some orthodontic patients can benefit from CBCT's capability to improve diagnosis and treatment planning. Appropriate use of CBCT by acquiring CBCT only when necessary has the potential to reduce ionizing radiation exposure to orthodontic patients. Generally, the risks of CBCT in orthodontics are outweighed by the benefits that CBCT scans provide in selected cases in which conventional radiographs cannot provide sufficient information necessary for diagnosis and treatment planning.

There is a strong consensus amongst position statements released by international organizations regarding CBCT in orthodontics, stating that CBCT is justified only when it brings a benefit to the patient or changes the outcome of the orthodontic treatment when compared with conventional imaging techniques. In these selected cases, the recommendation is to use the smallest possible FOV, with the lowest radiation exposure.

Therefore, CBCT can provide orthodontists with valuable diagnostic information, but its use should be case specific in which the clinician should be able to justify

the reason for CBCT acquisition. Prescribing CBCT regularly for all patients increases the collective dose for orthodontic patients and is not consistent with international guidelines for an appropriate use of ionizing radiation in orthodontics. Consequently, CBCT in orthodontics requires judicious and sound clinical judgement.

## References

1. De Vos, W.; Casselman, J.; Swennen, G.R. Cone-beam computerized tomography (CBCT) imaging of the oral and maxillofacial region: A systematic review of the literature. *Int. J. Oral Maxillofac. Surg.* 2009, *38*, 609–625. [CrossRef] [PubMed]
2. Leonardi, R. Cone-beam computed tomography and three-dimensional orthodontics. Where we are and future perspectives. *J. Orthod.* 2019, *46*, 45–48. [CrossRef] [PubMed]
3. Hall, E.; Giaccia, A. *Radiobiology for the Radiologist*, 6th ed.; Lippincott Williams and Wilkins: Philadelphia, PA, USA, 2006; pp. 135–153.
4. Buttke, T.M.; Proffit, W.R. Referring adult patients for orthodontic treatment. *J. Am. Dent. Assoc.* 1999, *130*, 73–79. [CrossRef] [PubMed]
5. Kleinerman, R.A. Cancer risks following diagnostic and therapeutic radiation exposure in children. *Pediatr. Radiol.* 2006, *36* (Suppl. 2), 121–125. [CrossRef]
6. Applegate, K.E.; Thomas, K. Pediatric CT—The challenge of dose records. *Pediatr. Radiol.* 2011, *41* (Suppl. 2), 523–527. [CrossRef]
7. Bulas, D.I.; Goske, M.J.; Applegate, K.E.; Wood, B.P. Image Gently: Why we should talk to parents about CT in children. *Am. J. Roentgenol.* 2009, *192*, 1176–1178. [CrossRef]
8. Brenner, D.J. Estimating cancer risks from pediatric CT: Going from the qualitative to the quantitative. *Pediatr. Radiol.* 2002, *32*, 228–221; discussion 242–244. [CrossRef]
9. Chodick, G.; Ronckers, C.M.; Shalev, V.; Ron, E. Excess lifetime cancer mortality risk attributable to radiation exposure from computed tomography examinations in children. *Isr. Med. Assoc. J. IMAJ* 2007, *9*, 584–587.
10. Isaacson, K.G.; Thom, A.R.; Atack, N.E.; Horner, K.; Whaites, E. *Orthodontic Radiographs: Guidelines for the Use of Radiographs in Clinical Orthodontics*, 4th ed.; British Orthodontic Society: London, UK, 2015.
11. Applegate, K.E.; Cost, N.G. Image Gently: A campaign to reduce children's and adolescents' risk for cancer during adulthood. *J. Adolesc. Health* 2013, *52*, S93–S97. [CrossRef]
12. Farman, A.G. ALARA still applies. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endod.* 2005, *100*, 395–397. [CrossRef]
13. Brenner, D.; Elliston, C.; Hall, E.; Berdon, W. Estimated risks of radiation-induced fatal cancer from pediatric CT. *Am. J. Roentgenol.* 2001, *176*, 289–296. [CrossRef] [PubMed]
14. Slovis, T.L. Children, computed tomography radiation dose, and the As Low As Reasonably Achievable (ALARA) concept. *Pediatrics* 2003, *112*, 971–972. [CrossRef] [PubMed]
15. European Commission. *European Guidelines on Radiation Protection in Dental Radiology*. Radiation Protection 136 Luxembourg: 2004. Available online: <https://ec.europa.eu/energy/sites/ener/files/documents/136.pdf>

- (accessed on 10 August 2019).
16. Hoffman, F.O.; Kocher, D.C.; Apostoaei, A.I. Beyond dose assessment: Using risk with full disclosure of uncertainty in public and scientific communication. *Health Phys.* 2011, *101*, 591–600. [CrossRef] [PubMed]
  17. McCollough, C.H.; Schueler, B.A. Calculation of effective dose. *Med. Phys.* 2000, *27*, 828–837. [CrossRef] [PubMed]
  18. Mozzo, P.; Procacci, C.; Tacconi, A.; Martini, P.T.; Andreis, I.A. A new volumetric CT machine for dental imaging based on the cone-beam technique: Preliminary results. *Eur. Radiol.* 1998, *8*, 1558–1564. [CrossRef] [PubMed]
  19. Mah, J.K.; Danforth, R.A.; Bumann, A.; Hatcher, D. Radiation absorbed in maxillofacial imaging with a new dental computed tomography device. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endod.* 2003, *96*, 508–513. [CrossRef]
  20. Bechara, B.; McMahan, C.A.; Geha, H.; Noujeim, M. Evaluation of a cone beam CT artefact reduction algorithm. *Dento Maxillo Facial Radiol.* 2012, *41*, 422–428. [CrossRef]
  21. Danforth, R.A.; Peck, J.; Hall, P. Cone beam volume tomography: An imaging option for diagnosis of complex mandibular third molar anatomical relationships. *J. Calif. Dent. Assoc.* 2003, *31*, 847–852.
  22. Sukovic, P. Cone beam computed tomography in craniofacial imaging. *Orthod. Craniofac. Res.* 2003, *6* (Suppl. 1), 31–36; discussion 179–182. [CrossRef]
  23. Silva, M.A.; Wolf, U.; Heinicke, F.; Bumann, A.; Visser, H.; Hirsch, E. Cone-beam computed tomography for routine orthodontic treatment planning: A radiation dose evaluation. *Am. J. Orthod. Dentofac. Orthop.* 2008, *133*, 640.e1–640.e5. [CrossRef]
  24. Bornstein, M.M.; Scarfe, W.C.; Vaughn, V.M.; Jacobs, R. Cone beam computed tomography in implant dentistry: A systematic review focusing on guidelines, indications, and radiation dose risks. *Int. J. Oral Maxillofac. Implant.* 2014, *29*, 55–77. [CrossRef] [PubMed]
  25. Roberts, J.A.; Drage, N.A.; Davies, J.; Thomas, D.W. Effective dose from cone beam CT examinations in dentistry. *Br. J. Radiol.* 2009, *82*, 35–40. [CrossRef] [PubMed]
  26. McCollough, C.H.; Bushberg, J.T.; Fletcher, J.G.; Eckel, L.J. Answers to Common Questions About the Use and Safety of CT Scans. *Mayo Clin. Proc.* 2015, *90*, 1380–1392. [CrossRef] [PubMed]
  27. Hedesiu, M.; Marcu, M.; Salmon, B.; Pauwels, R.; Oenning, A.C.; Almasan, O.; Roman, R.; Baciut, M.; Jacobs, R. Irradiation provided by dental radiological procedures in a pediatric population. *Eur. J. Radiol.* 2018, *103*, 112–117. [CrossRef] [PubMed]
  28. Pauwels, R.; Beinsberger, J.; Collaert, B.; Theodorakou, C.; Rogers, J.; Walker, A.; Cockmartin, L.; Bosmans, H.; Jacobs, R.; Bogaerts, R.; et al. Effective dose range for dental cone beam computed tomography scanners. *Eur. J. Radiol.* 2012, *81*, 267–271. [CrossRef] [PubMed]
  29. Ludlow, J.B.; Davies-Ludlow, L.E.; Brooks, S.L.; Howerton, W.B. Dosimetry of 3 CBCT devices for oral and maxillofacial radiology: CB Mercuray, NewTom 3G and i-CAT. *Dento Maxillo Facial Radiol.* 2006, *35*, 219–226. [CrossRef] [PubMed]
  30. Yeung, A.W.K.; Jacobs, R.; Bornstein, M.M. Novel low-dose protocols using cone beam computed tomography in dental medicine: A review focusing on

- indications, limitations, and future possibilities. *Clin. Oral Investig.* 2019, 23, 2573–2581. [CrossRef] [PubMed]
31. Mah, J.K.; Danforth, R.A. Comparative direct dosimetry of cone-beam computed tomography using reduced basis projections. *J. Clin. Orthod. JCO* 2018, 52, 173–179.
  32. Oenning, A.C.; Pauwels, R.; Stratis, A.; De Faria Vasconcelos, K.; Tijskens, E.; De Grauwe, A.; Jacobs, R.; Salmon, B. Halve the dose while maintaining image quality in paediatric Cone Beam CT. *Sci. Rep.* 2019, 9, 5521. [CrossRef]
  33. Pauwels, R.; Jacobs, R.; Bogaerts, R.; Bosmans, H.; Panmekiate, S. Determination of size-specific exposure settings in dental cone-beam CT. *Eur. Radiol.* 2017, 27, 279–285. [CrossRef]
  34. Ludlow, J.B.; Walker, C. Assessment of phantom dosimetry and image quality of i-CAT FLX cone-beam computed tomography. *Am. J. Orthod. Dentofac. Orthop.* 2013, 144, 802–817. [CrossRef] [PubMed]
  35. Kapila, S.; Conley, R.S.; Harrell, W.E., Jr. The current status of cone beam computed tomography imaging in orthodontics. *Dento Maxillo Facial Radiol.* 2011, 40, 24–34. [CrossRef] [PubMed]
  36. Oenning, A.C.; Jacobs, R.; Pauwels, R.; Stratis, A.; Hedesiu, M.; Salmon, B. Cone-beam CT in paediatric dentistry: DIMITRA project position statement. *Pediatr. Radiol.* 2018, 48, 308–316. [CrossRef] [PubMed]
  37. Ludlow, J.B.; Timothy, R.; Walker, C.; Hunter, R.; Benavides, E.; Samuelson, D.B. Correction to Effective dose of dental CBCT—A meta analysis of published data and additional data for nine CBCT units. *Dento Maxillo Facial Radiol.* 2015, 44, 20159003. [CrossRef] [PubMed]
  38. Jacobs, R.; Pauwels, R.; Scarfe, W.C.; De Cock, C.; Dula, K.; Willems, G.; Verdonck, A.; Politis, C. Pediatric cleft palate patients show a 3-to 5-fold increase in cumulative radiation exposure from dental radiology compared with an age-and gender-matched population: A retrospective cohort study. *Clin. Oral Investig.* 2018, 22, 1783–1793. [CrossRef] [PubMed]
  39. Marcu, M.; Hedesiu, M.; Salmon, B.; Pauwels, R.; Stratis, A.; Oenning, A.C.C.; Cohen, M.E.; Jacobs, R.; Baciut, M.; Roman, R.; et al. Estimation of the radiation dose for pediatric CBCT indications: A prospective study on ProMax3D. *Int. J. Paediatr. Dent.* 2018, 28, 300–309. [CrossRef]
  40. Pauwels, R.; Jacobs, R.; Bogaerts, R.; Bosmans, H.; Panmekiate, S. Reduction of scatter-induced image noise in cone beam computed tomography: Effect of field of view size and position. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol.* 2016, 121, 188–195. [CrossRef]
  41. Newaz, Z.A.; Barghan, S.; Katkar, R.A.; Bennett, J.A.; Nair, M.K. Incidental findings of skull-base abnormalities in cone-beam computed tomography scans with consultation by maxillofacial radiologists. *Am. J. Orthod. Dentofac. Orthop.* 2015, 147, 127–131. [CrossRef]
  42. Okano, T.; Harata, Y.; Sugihara, Y.; Sakaino, R.; Tsuchida, R.; Iwai, K.; Seki, K.; Araki, K. Absorbed and effective doses from cone beam volumetric imaging for implant planning. *Dento Maxillo Facial Radiol.* 2009, 38, 79–85. [CrossRef]
  43. Garcia Silva, M.A.; Wolf, U.; Heinicke, F.; Grundler, K.; Visser, H.; Hirsch, E. Effective dosages for recording Veraviewepocs dental panoramic images: Analog film, digital, and panoramic scout for CBCT. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endod.* 2008, 106, 571–577. [CrossRef]

44. Gijbels, F.; Jacobs, R.; Bogaerts, R.; Debaveye, D.; Verlinden, S.; Sanderink, G. Dosimetry of digital panoramic imaging. Part I: Patient exposure. *Dento Maxillo Facial Radiol.* 2005, *34*, 145–149. [CrossRef] [PubMed]
45. Lecomber, A.R.; Yoneyama, Y.; Lovelock, D.J.; Hosoi, T.; Adams, A.M. Comparison of patient dose from imaging protocols for dental implant planning using conventional radiography and computed tomography. *Dento Maxillo Facial Radiol.* 2001, *30*, 255–259. [CrossRef]
46. Theodorakou, C.; Walker, A.; Horner, K.; Pauwels, R.; Bogaerts, R.; Jacobs, R. Estimation of paediatric organ and effective doses from dental cone beam CT using anthropomorphic phantoms. *Br. J. Radiol.* 2012, *85*, 153–160. [CrossRef] [PubMed]
47. Gavala, S.; Donta, C.; Tsiklakis, K.; Boziari, A.; Kamenopoulou, V.; Stamatakis, H.C. Radiation dose reduction in direct digital panoramic radiography. *Eur. J. Radiol.* 2009, *71*, 42–48. [CrossRef] [PubMed]
48. Gijbels, F.; Sanderink, G.; Wyatt, J.; Van Dam, J.; Nowak, B.; Jacobs, R. Radiation doses of indirect and direct digital cephalometric radiography. *Br. Dent. J.* 2004, *197*, 149–152; discussion 140. [CrossRef] [PubMed]
49. Ludlow, J.B.; Davies-Ludlow, L.E.; Brooks, S.L. Dosimetry of two extraoral direct digital imaging devices: NewTom cone beam CT and Orthophos Plus DS panoramic unit. *Dento Maxillo Facial Radiol.* 2003, *32*, 229–234. [CrossRef] [PubMed]
50. Danforth, R.A.; Dus, I.; Mah, J. 3-D volume imaging for dentistry: A new dimension. *J. Calif. Dent. Assoc.* 2003, *31*, 817–823.
51. Signorelli, L.; Patcas, R.; Peltomaki, T.; Schatzle, M. Radiation dose of cone-beam computed tomography compared to conventional radiographs in orthodontics. *J. Orofac. Orthop.* 2016, *77*, 9–15. [CrossRef]
52. Ludlow, J.B.; Ivanovic, M. Comparative dosimetry of dental CBCT devices and 64-slice CT for oral and maxillofacial radiology. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endod.* 2008, *106*, 106–114. [CrossRef]
53. Tsiklakis, K.; Donta, C.; Gavala, S.; Karayianni, K.; Kamenopoulou, V.; Hourdakakis, C.J. Dose reduction in maxillofacial imaging using low dose Cone Beam CT. *Eur. J. Radiol.* 2005, *56*, 413–417. [CrossRef]
54. Ludlow, J.B.; Davies-Ludlow, L.E.; White, S.C. Patient risk related to common dental radiographic examinations: The impact of 2007 International Commission on Radiological Protection recommendations regarding dose calculation. *J. Am. Dent. Assoc.* 2008, *139*, 1237–1243. [CrossRef] [PubMed]
55. Hans, M.G.; Palomo, J.M.; Valiathan, M. History of imaging in orthodontics from Broadbent to cone-beam computed tomography. *Am. J. Orthod. Dentofac. Orthop.* 2015, *148*, 914–921. [CrossRef] [PubMed]
56. Huda, W.; Atherton, J.V.; Ware, D.E.; Cumming, W.A. An approach for the estimation of effective radiation dose at CT in pediatric patients. *Radiology* 1997, *203*, 417–422. [CrossRef] [PubMed]
57. White, S.C.; Scarfe, W.C.; Schulze, R.K.; Lurie, A.G.; Douglass, J.M.; Farman, A.G.; Law, C.S.; Levin, M.D.; Sauer, R.A.; Valachovic, R.W.; et al. The Image Gently in Dentistry campaign: Promotion of responsible use of maxillofacial radiology in dentistry for children. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol.* 2014, *118*, 257–261. [CrossRef] [PubMed]
58. Mettler, F.A., Jr.; Thomadsen, B.R.; Bhargavan, M.; Gilley, D.B.; Gray, J.E.; Lipoti, J.A.; McCrohan, J.; Yoshizumi, T.T.; Mahesh, M. Medical

- radiation exposure in the U.S. in 2006: Preliminary results. *Health Phys.* 2008, 95, 502–507. [CrossRef] [PubMed]
59. Fazel, R.; Krumholz, H.M.; Wang, Y.; Ross, J.S.; Chen, J.; Ting, H.H.; Shah, N.D.; Nasir, K.; Einstein, A.J.; Nallamothu, B.K. Exposure to low-dose ionizing radiation from medical imaging procedures. *N. Engl. J. Med.* 2009, 361, 849–857. [CrossRef] [PubMed]
  60. Brenner, D.J.; Doll, R.; Goodhead, D.T.; Hall, E.J.; Land, C.E.; Little, J.B.; Lubin, J.H.; Preston, D.L.; Preston, R.J.; Puskin, J.S.; et al. Cancer risks attributable to low doses of ionizing radiation: Assessing what we really know. *Proc. Natl. Acad. Sci. USA* 2003, 100, 13761–13766. [CrossRef] [PubMed]
  61. Yeh, J.K.; Chen, C.H. Estimated radiation risk of cancer from dental cone-beam computed tomography imaging in orthodontics patients. *BMC Oral Health* 2018, 18, 131. [CrossRef] [PubMed]
  62. Pauwels, R. Cone beam CT for dental and maxillofacial imaging: Dose matters. *Radiat. Prot. Dosim.* 2015, 165, 156–161. [CrossRef]
  63. American Dental Association Council on Scientific Affairs. The use of dental radiographs: Update and recommendations. *J. Am. Dent. Assoc.* 2006, 137, 1304–1312.
  64. Kamburoglu, K. Use of dentomaxillofacial cone beam computed tomography in dentistry. *World J. Radiol.* 2015, 7, 128–130. [CrossRef] [PubMed]
  65. Weiss, R., 2nd; Read-Fuller, A. Cone Beam Computed Tomography in Oral and Maxillofacial Surgery: An Evidence-Based Review. *Dent. J.* 2019, 7, 52. [CrossRef] [PubMed]
  66. Hsieh, J.; Molthen, R.C.; Dawson, C.A.; Johnson, R.H. An iterative approach to the beam hardening correction in cone beam CT. *Med. Phys.* 2000, 27, 23–29. [CrossRef] [PubMed]
  67. Pauwels, R.; Stamatakis, H.; Bosmans, H.; Bogaerts, R.; Jacobs, R.; Horner, K.; Tsiklakis, K. Quantification of metal artifacts on cone beam computed tomography images. *Clin. Oral Implant. Res.* 2013, 24 (Suppl. A100), 94–99. [CrossRef]
  68. Hirschinger, V.; Hanke, S.; Hirschfelder, U.; Hofmann, E. Artifacts in orthodontic bracket systems in cone-beam computed tomography and multislice computed tomography. *J. Orofac. Orthop.* 2015, 76, 152–163. [CrossRef] [PubMed]
  69. Endo, M.; Tsunoo, T.; Nakamori, N.; Yoshida, K. Effect of scattered radiation on image noise in cone beam CT. *Med. Phys.* 2001, 28, 469–474. [CrossRef] [PubMed]
  70. Coskun, I.; Kaya, B. Cone Beam Computed Tomography in Orthodontics. *Turk. J. Orthod.* 2018, 31, 55–61. [CrossRef] [PubMed]
  71. Alqerban, A.; Jacobs, R.; Fieuws, S.; Nackaerts, O.; Willems, G. Comparison of 6 cone-beam computed tomography systems for image quality and detection of simulated canine impaction-induced external root resorption in maxillary lateral incisors. *Am. J. Orthod. Dentofac. Orthop.* 2011, 140, e129–e139. [CrossRef] [PubMed]
  72. Mischkowski, R.A.; Pulsfort, R.; Ritter, L.; Neugebauer, J.; Brochhagen, H.G.; Keeve, E.; Zoller, J.E. Geometric accuracy of a newly developed cone-beam device for maxillofacial imaging. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endod.* 2007, 104, 551–559. [CrossRef]
  73. Moreira, C.R.; Sales, M.A.; Lopes, P.M.; Cavalcanti, M.G. Assessment of linear

- and angular measurements on three-dimensional cone-beam computed tomographic images. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endod.* 2009, *108*, 430–436. [CrossRef]
74. El, H.; Palomo, J.M. Measuring the airway in 3 dimensions: A reliability and accuracy study. *Am. J. Orthod. Dentofac. Orthop.* 2010, *137*, e51–e59; discussion S50–S52. [CrossRef]
  75. Al-Ekrish, A.A.; Ekram, M. A comparative study of the accuracy and reliability of multidetector computed tomography and cone beam computed tomography in the assessment of dental implant site dimensions. *Dento Maxillo Facial Radiol.* 2011, *40*, 67–75. [CrossRef] [PubMed]
  76. Ganguly, R.; Ruprecht, A.; Vincent, S.; Hellstein, J.; Timmons, S.; Qian, F. Accuracy of linear measurement in the Galileos cone beam computed tomography under simulated clinical conditions. *Dento Maxillo Facial Radiol.* 2011, *40*, 299–305. [CrossRef] [PubMed]
  77. Gribel, B.F.; Gribel, M.N.; Frazao, D.C.; McNamara, J.A., Jr.; Manzi, F.R. Accuracy and reliability of craniometric measurements on lateral cephalometry and 3D measurements on CBCT scans. *Angle Orthod.* 2011, *81*, 26–35. [CrossRef] [PubMed]
  78. Timock, A.M.; Cook, V.; McDonald, T.; Leo, M.C.; Crowe, J.; Benninger, B.L.; Covell, D.A., Jr. Accuracy and reliability of buccal bone height and thickness measurements from cone-beam computed tomography imaging. *Am. J. Orthod. Dentofac. Orthop.* 2011, *140*, 734–744. [CrossRef] [PubMed]
  79. Razavi, T.; Palmer, R.M.; Davies, J.; Wilson, R.; Palmer, P.J. Accuracy of measuring the cortical bone thickness adjacent to dental implants using cone beam computed tomography. *Clin. Oral Implant. Res.* 2010, *21*, 718–725. [CrossRef] [PubMed]
  80. Horner, K.; Islam, M.; Flygare, L.; Tsiklakis, K.; Whaites, E. Basic principles for use of dental cone beam computed tomography: Consensus guidelines of the European Academy of Dental and Maxillofacial Radiology. *Dento Maxillo Facial Radiol.* 2009, *38*, 187–195. [CrossRef]
  81. Brown, J.; Jacobs, R.; Levring Jaghagen, E.; Lindh, C.; Baksi, G.; Schulze, D.; Schulze, R. Basic training requirements for the use of dental CBCT by dentists: A position paper prepared by the European Academy of DentoMaxilloFacial Radiology. *Dento Maxillo Facial Radiol.* 2014, *43*, 20130291. [CrossRef]
  82. Jerrold, L. Litigation, legislation, and ethics. Liability regarding computerized axial tomography scans. *Am. J. Orthod. Dentofac. Orthop.* 2007, *132*, 122–124. [CrossRef]
  83. Turpin, D.L. Befriend your oral and maxillofacial radiologist. *Am. J. Orthod. Dentofac. Orthop.* 2007, *131*, 697. [CrossRef]
  84. Friedland, B. Medicolegal issues related to cone beam CT. *Semin. Orthod.* 2009, *15*, 77–84. [CrossRef]
  85. American Academy of Oral and Maxillofacial Radiology. Clinical recommendations regarding use of cone beam computed tomography in orthodontics. [corrected]. Position statement by the American Academy of Oral and Maxillofacial Radiology. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol.* 2013, *116*, 238–257. [CrossRef] [PubMed]
  86. Zinman, E.J.; White, S.C.; Tetradis, S. Legal considerations in the use of cone beam computer tomography imaging. *J. Calif. Dent. Assoc.* 2010, *38*, 49–56. [PubMed]

87. Abdelkarim, A.; Jerrold, L. Clinical considerations and potential liability associated with the use of ionizing radiation in orthodontics. *Am. J. Orthod. Dentofac. Orthop.* 2018, *154*, 15–25. [CrossRef] [PubMed]
88. Lombardo, L.; Arreghini, A.; Guarneri, M.P.; Lauritano, D.; Nardone, M.; Siciliani, G. Unexpected artefacts and occult pathologies under CBCT. *Oral Implantol.* 2017, *10*, 97–104. [CrossRef] [PubMed]
89. Curley, A.; Hatcher, D.C. Cone beam CT—Anatomic assessment and legal issues: The new standards of care. *J. Calif. Dent. Assoc.* 2009, *37*, 653–662. [PubMed]
90. Bruks, A.; Enberg, K.; Nordqvist, I.; Hansson, A.S.; Jansson, L.; Svenson, B. Radiographic examinations as an aid to orthodontic diagnosis and treatment planning. *Swed. Dent. J.* 1999, *23*, 77–85. [PubMed]
91. Atchison, K.A.; Luke, L.S.; White, S.C. An algorithm for ordering pretreatment orthodontic radiographs. *Am. J. Orthod. Dentofac. Orthop.* 1992, *102*, 29–44. [CrossRef]
92. Dindaroglu, F.; Yetkiner, E. Cone Beam Computed Tomography in Orthodontics. *Turk. J. Orthod.* 2016, *29*, 16–21. [CrossRef]
93. American Dental Association Council on Scientific Affairs. The use of cone-beam computed tomography in dentistry: An advisory statement from the American Dental Association Council on Scientific Affairs. *J. Am. Dent. Assoc.* 2012, *143*, 899–902.
94. De Grauwe, A.; Ayaz, I.; Shujaat, S.; Dimitrov, S.; Gbadegbegnon, L.; Vande Vannet, B.; Jacobs, R. CBCT in orthodontics: A systematic review on justification of CBCT in a paediatric population prior to orthodontic treatment. *Eur. J. Orthod.* 2018. [CrossRef] [PubMed]
95. Chinem, L.A.; Vilella Bde, S.; Mauricio, C.L.; Canevaro, L.V.; Deluiz, L.F.; Vilella Ode, V. Digital orthodontic radiographic set versus cone-beam computed tomography: An evaluation of the effective dose. *Dent. Press J. Orthod.* 2016, *21*, 66–72. [CrossRef] [PubMed]
96. Dula, K.; Benic, G.I.; Bornstein, M.; Dagassan-Berndt, D.; Filippi, A.; Hicklin, S.; Kissling-Jeger, F.; Luebbers, H.T.; Sculean, A.; Sequeira-Byron, P.; et al. SADMFR Guidelines for the Use of Cone-Beam Computed Tomography/Digital Volume Tomography. *Swiss Dent. J.* 2015, *125*, 945–953. [PubMed]
97. Greco, P.M. Ethics in orthodontics. Let the truth be known. *Am. J. Orthod. Dentofac. Orthop.* 2013, *144*, 788–789. [CrossRef] [PubMed]
98. Alqerban, A.; Willems, G.; Bernaerts, C.; Vangastel, J.; Politis, C.; Jacobs, R. Orthodontic treatment planning for impacted maxillary canines using conventional records versus 3D CBCT. *Eur. J. Orthod.* 2014, *36*, 698–707. [CrossRef] [PubMed]
99. Hujoel, P.P.; Aps, J.K.; Bollen, A.M. What are the cancer risks from dental computed tomography? *J. Dent. Res.* 2015, *94*, 7–9. [CrossRef] [PubMed]
100. Halazonetis, D.J. Cone-beam computed tomography is not the imaging technique of choice for comprehensive orthodontic assessment. *Am. J. Orthod. Dentofac. Orthop.* 2012, *141*, 407. [CrossRef]
101. Kokich, V.G. Cone-beam computed tomography: Have we identified the orthodontic benefits? *Am. J. Orthod. Dentofac. Orthop.* 2010, *137*, S16. [CrossRef]
102. Van Vlijmen, O.J.; Kuijpers, M.A.; Berge, S.J.; Schols, J.G.; Maal, T.J.; Breuning, H.; Kuijpers-Jagtman, A.M. Evidence supporting the use of cone-



- beam computed tomography in orthodontics. *J. Am. Dent. Assoc.* 2012, 143, 241–252. [CrossRef]
103. Tadinada, A.; Marczak, A.; Yadav, S.; Mukherjee, P.M. Applications of Cone Beam Computed Tomography in Orthodontics: A Review. *Turk. J. Orthod.* 2016, 29, 73–79. [CrossRef]
  104. Merrett, S.J.; Drage, N.A.; Durning, P. Cone beam computed tomography: A useful tool in orthodontic diagnosis and treatment planning. *J. Orthod.* 2009, 36, 202–210. [CrossRef] [PubMed]
  105. Maverna, R.; Gracco, A. Different diagnostic tools for the localization of impacted maxillary canines: Clinical considerations. *Prog. Orthod.* 2007, 8, 28–44. [PubMed]
  106. Liu, D.G.; Zhang, W.L.; Zhang, Z.Y.; Wu, Y.T.; Ma, X.C. Localization of impacted maxillary canines and observation of adjacent incisor resorption with cone-beam computed tomography. *Oral Surg. Oral Med. Oral Pathol. Oral Radiol. Endod.* 2008, 105, 91–98. [CrossRef] [PubMed]